



Catalog No. MAL10025-10 10 Tests
Catalog No. MAL10025 25 Tests

RAPID TEST FOR MALARIA Pan/Pf

For in vitro diagnostic use

INTRODUCTION

RAPID TEST FOR MALARIA Pan / Pf is a rapid qualitative, two site sandwich immunoassay, utilizing whole blood for the detection of *P.falciparum* specific histidine rich protein-2 (Pf HRP-2) and pan specific pLDH. The test may also be used for differentiation of *P. falciparum* and other malarial species and for the follow up of antimalarial therapy, in whole blood samples.

SUMMARY

Four species of the Plasmodium parasites are responsible for malaria infections in human viz. *P. falciparum*, *P.vivax*, *P.ovale* and *P.malariae*. Of these *P. falciparum* and *P. vivax* are the most prevalent. Early detection and differentiation of malaria is of paramount importance due to incidence of cerebral malaria and drug resistance associated with falciparum malaria and due to the morbidity associated with the other malarial forms.

RAPID TEST FOR MALARIA Pan / Pf detects the presence of pan malaria specific pLDH released from the parasitised blood cells, for the detection of all malarial parasites. Whereas, for the detection of *P. falciparum* malaria, **RAPID TEST FOR MALARIA Pan / Pf** utilizes the detection of *P.falciparum* specific histidine rich protein-2 (Pf HRP-2), which is a water soluble protein that is released from parasitized erythrocytes of infected individuals.

In the absence of *P.falciparum* specific Pf HRP-2, the presence of the pan malaria specific band points to the presence of other malarial species; viz.; *P.vivax*, *P.ovale* or *P.malariae*. Speciation is done and results inferred in the context of prevalence rates of the malarial species prevalent in the particular region.

Since pLDH is a product of viable parasites, the pan band may also be used to monitor success of antimalarial therapy.

PRINCIPLE

RAPID TEST FOR MALARIA Pan / Pf utilizes the principle of immunochromatography. As the test sample flows through the membrane assembly of the device after addition of the clearing buffer, the colored monoclonal anti HRP-2 specific / anti pan specific -colloidal gold conjugate antibodies complexes the proteins in the lysed sample. This complex moves further on the membrane to the test region where it is immobilized by the monoclonal anti HRP-2 / anti pan specific antibody coated on the membrane leading to formation of pink-purple colored band/s which confirms a positive test result. While both the bands will appear at the test region in falciparum positive samples, only one band would appear in non-falciparum malaria positive samples. Absence of this colored band/s in the test region indicates a negative test result.

The unreacted conjugate along with the rabbit globulin colloidal gold conjugate and unbound complex if any, move further on the membrane and are subsequently immobilized by anti-rabbit antibodies coated on the membrane at the control region, forming a pink-purple band. This control band serves to validate the test performance.

REAGENTS AND MATERIAL SUPPLIED

RAPID TEST FOR MALARIA Pan / Pf kit contains:

A. Individual pouches, each containing:

1. Test Device: Membrane assembly predisposed with monoclonal anti HRP-2 - colloidal gold conjugate, monoclonal anti pan specific pLDH -colloidal gold conjugate, rabbit globulin-colloidal gold conjugate and monoclonal anti Pf HRP-2 antibody, monoclonal anti pan specific pLDH antibody and anti-rabbit antibody at the respective regions.
 2. Desiccant pouch.
 3. 5 µl sample loop.
- B. Clearing buffer in a dropper bottle.
C. Package insert.

OPTIONAL MATERIAL REQUIRED

Calibrated micro pipettes capable of delivering 5µl of sample accurately.

STORAGE AND STABILITY

The test kit may be stored between 4-30°C till the duration of the shelf life as indicated on the pouch / carton. DO NOT FREEZE.

NOTE

Read the instructions carefully before performing the test. For in vitro diagnostic use only. NOT FOR MEDICINAL USE. Do not use beyond expiry date. Do not inter mix reagents from different lots. Handle all specimens as potentially infectious. Follow standard biosafety guidelines for handling and disposal of potentially infective material.

SPECIMEN COLLECTION AND PREPARATION

Fresh anti coagulated whole blood should be used as a test sample and EDTA or Heparin or Oxalate can be used as suitable anticoagulant. The specimen should be collected in a clean glass or plastic container. If immediate testing is not possible then the specimen may be stored at 2–8°C for up to 72 hours before testing. Clotted or

contaminated blood samples should not be used for performing the test. Fresh blood from finger prick / puncture may also be used as a test specimen.

TEST PROCEDURE

1. Bring the **RAPID TEST FOR MALARIA Pan / Pf** kit components to room temperature before testing.
2. Open the pouch and retrieve the device, sample loop and desiccant. Check the color of the desiccant. It should be blue. If it has turned colorless or pink, discard the device and use another device. *Once opened, the device must be used immediately.*
3. Tighten the vial cap of the clearing buffer provided with the kit in the clockwise direction to pierce the dropper bottle nozzle.
4. Evenly mix the anti coagulated blood sample by gentle swirling. Dip the sample loop into the sample. Ensuring that a loop full of blood is retrieved, blot the blood so collected on to the sample pad in the sample port 'A'. (This delivers approximately 5 µl of the whole blood specimen)

OR

In case finger prick blood is being used, touch the sample loop to the blood on the finger prick. Ensuring that a loop full of blood is retrieved, immediately blot the specimen on to the sample pad in the sample port 'A' (Care should be taken that the blood sample has not clotted and the transfer to the sample pad is immediate).

OR

Alternatively, 5 µl of the anticoagulated or finger prick specimen may be delivered to the sample pad in the sample port 'A' using a micropipette.


NOTE: Ensure the blood from the sample applicator pipette has been completely taken up by the sample pad.

5. Dispense four drops of the clearing buffer into port 'B', by holding the plastic dropper bottle vertically.

6. Read the results at the end of 15 minutes. However, if the background of the result window has not cleared within this time, wait for another 15 minutes before reading the results as follows:

NEGATIVE for malaria:

Only one pink-purple band appears in the control window 'C'.




POSITIVE for malaria:


***P. falciparum* or mixed infection:** In addition to the control band, two pink-purple bands appear at regions 'Pf' and 'Pan' in the test window 'T'.



Other species (Non falciparum): In addition to the control band, one pink-purple band appears only at region 'Pan' in the test window 'T'.



On rare occasions a Pf band is present but the Pan band is negative, this may indicate post treatment malaria or in some cases untreated malaria. The test should be repeated in 2 days.



7. The test should be considered invalid if no bands appear on the device. Repeat the test with a new device ensuring that the test procedure has been followed accurately.

LIMITATIONS OF THE TEST

1. As with all diagnostic tests, the test result must always be correlated with clinical findings
2. The results of test are to be interpreted within the epidemiological, clinical and therapeutic

context. When it seems indicated, the parasitological techniques of reference should be considered (microscopic examination of the thick smear and thin blood films).

3. Any modification to the above procedure and / or use of other reagents will invalidate the test procedure.
4. The device and buffer of different lots must not be mixed and used.
5. In case of mixed infection (*P. falciparum* with other malarial species), both, Pf and pan malaria band will be positive. Hence differentiation of infection due to *P. vivax*, *P. ovale* or *P. malariae* cannot be done.
6. While monitoring therapy, using the pan band, if the reaction of the test remains positive with the same intensity after 5-10 days, post treatment, the possibility of a resistant strain of malaria has to be considered.
7. Usually, the pan band turn negative after successful anti malarial therapy. However, since treatment duration and medication used affect the clearance of parasites, the test should be repeated after 5-10 days of start of treatment.
8. In *P. falciparum* malaria infection, HRP-2 is not secreted in the gametogony stage. Hence, in "Carriers", the HRP-2 band may be absent.
9. HRP-2 levels, post treatment persists upto 15 days; the pan band can be used to monitor success of therapy, in *P. falciparum* malaria cases.
10. In a few cases, where the HRP-2 band is positive and the pan malaria band is negative, it may indicate a case of post treatment malaria. However, such a reaction pattern may also be obtained in a few cases of untreated malaria. Retesting after 2 days is advised, in such cases.
11. Most blood samples clear within running the test. However, in a few fresh samples and especially in stored samples, the background clearance may be delayed for 15-20 minutes more. In such cases it is strongly recommended to extend the reading time by another 15 minutes so that the results can be interpreted against a clear background.

PERFORMANCE CHARACTERISTICS

Sample	Total No. of samples tested	RAPID TEST FOR MALARIA		Sensitivity (%)	Specificity (%)
		(+)	(-)		
<i>P. falciparum</i> positive	16	16	0	100%	-
<i>P. vivax</i> positive	25	25	0	100%	-
Malaria negative	210	0	210	-	100%

BIBLIOGRAPHY

- Howard, R.J., et al, 1986: Secretion of a Malarial Histidine-rich Protein (Pf. HRP II) from *Plasmodium falciparum*-infected Erythrocytes. *J. Cell Biol.*, 103, 1269-1277.
- Rock, E.P., et al, 1987: Comparative Analysis of the *Plasmodium falciparum* Histidine-Rich Proteins HRP-I, HRP-II, and HRP-III in Malaria Parasities of Diverse Origin. *Parasitol.*, 95, 209-227.
- Parra, M.E., et al, 1991: Identification of *Plasmodium falciparum* Histidine-Rich Protein 2 in the Plasma of Humans with Malaria. *J. Clin. Microbiol.*, 29, 1629-1634.

In an in house study, a panel of 251 samples whose results were earlier confirmed with

- Rodriguez-Del Valle, M., et al, 1991: Detection of Antigens and Antibodies in the Urine of Humans with *Plasmodium falciparum* Malaria. *J. Clin. Microbiol.*, 29, 1236-1242.
- Makler, M. T., et. al.(1993) Parasite lactate assay as an assay for *Plasmodium falciparum* drug sensitivity. *Am. J. Trop. Med. Hyg.* 48(6), 739-741.
- Piper, R. C., et. al., (1999) Immuno-capture diagnostic assays for malaria utilizing *Plasmodium* Lactate Dehydrogenase (pLDH) *Am. J. Trop. Med.Hyg.* 60(1) 109-118.
- Srinivasan. S., et. al.,(2000) Comparison of blood – film microscopy, The OptiMAL dipstick, Rhodamine- 123 fluorescence staining and PCR for monitoring antimalarial treatment. *Annals of Tropical Medicine and Parasitology*, 94(3) 227-232.
- Hunte-Cooke A., et. al., (1999) Comparison of a Parasite Lactate Dehydrogenase-based Immunochromatographic. Antigen Detection assay (OptiMAL®) with Microscopy for the Detection of Malaria Parasites in Human Blood Samples. *Am J.Trop Med* 60(2). 173-176.
- John, S. M., et. al.,(1998) Evaluation of OptiMAL,a dipstick test for the diagnosis of malaria. *Ann. Trop. Med. Parasitol.*, 92, 621-622.

microscopy were tested with **parascreen**. The results obtained are as follows:

- Quintana M., et. al.,(1998) Malaria diagnosis by dipstick assay in a Honduran Population with coendemic *Plasmodium falciparum* and *Plasmodium vivax*. *Am. J. Trop. Med. Hyg.* 59(6) 868-871.
- Palmer, C. J.,((1998) Evaluation of OptiMal test for rapid diagnosis of *Plasmodium vivax* and *Plasmodium falciparum* . *J. Clin Microbiol* . 36(1) 203-206.
- Moody A., et. al (2000) Performance of the OptiMAL® malaria antigen capture dipstick for malaria diagnosis and treatment monitoring. *British Journal of Hematology*, 109, 1-5 .



Nova Century Scientific
 5022 South Service Road
 Burlington, Ontario L7L 5Y7
 Tel: 1-800-615-5072
 Fax: 1-800-639-9006
www.novacentury.com